



Illinois Department of Public Health
Emergency Medical Systems

Ambulance Inspection Form

Provider: License#:

Garage Address:

VIN: Level of Care: Local ID: EMS System#

Inspection Type: Inspected By:

Inspection Date: Inspection Status: Action:

- [1] Wheeled multi-level cot w/3 sets if straps + over shoulder straps
[2] 3-Point fastener for cot
[3] Cot fits securely in fastener
[4] Secondary stretcher w/3 sets of straps

On-board oxygen equipment (5-8)

- [5] Cylinder is not empty --volume:
[6] Delivery tubes
[7] 1 each A, C, I masks
[8] 3 each nasal cannula

Portable Oxygen Equipment(9-15)

- [9] Full primary cylinder minimum size = 'D' --volume:
[10] Dial flowmeter/regulator for 15 lpm
[11] Delivery tubes
[12] 1 each A, C, I masks
[13] Full spare cylinder
[14] Quick release, crash-stable, mounting bracket for cylinders

OR

- [15] Quick release mounting bracket if stored in a secure compartment

On board Suction Equipment(16-21)

- [16] Obtains 300mm within 4 seconds of clamping tubing
[17] Has 10' transparent non-kinking tubing / 1/4" inside diameter
[18] Can adjust vacuum levels
[19] Collection bottle holds 1000ml
[20] 3 sterile, single-use catheters, one each size: S/6-8F - M/10-12F - L/14-18F
[21] 3 Sterile, semi-rigid pharyngeal suction tips

Portable Suction Equipment(22-31)

- [22] Obtains 300mm within 4 seconds of clamping tubing
[23] Tubing is translucent, non-kinking or collapsing
[24] Tubing is 1/4" inside diameter
[25] 3 Sterile, single-use catheters, one each size: S/6-8F - M/10-12F - L14-18F
[26] 3 Sterile, semi-rigid pharyngeal suction tips
[27] Gas powered portable unit meets above requirements
[28] Capable of charging from vehicle 12 volt system and 115 volt AC
[29] Operates from internal rechargeable battery supply
[30] Operates for 20 continuous minutes (perform if battery sounds weak)

OR

- [31] Manually operated suction device (IDPH approved)

Medical Equipment(32-51)

- [32] Adult squeeze bag-valve-mask with adult and child masks
[33] Child squeeze bag-valve-mask with child and infant masks
[34] Traction splint, adult
[35] Traction splint, pediatric
[36] Adult blood pressure cuff/gauge
[37] Child blood pressure cuff/gauge
[38] Infant blood pressure cuff/gauge
[39] Stethoscope (2)
[40] Long spine board (adult) with 3 sets of straps
[41] Short spine board (32" x 16" min) w/2 9' straps or vest type (wrap around) device (optional)
[42] Airway, Oropharyngeal, A, C, I
[43] Airway, Nasopharyngeal with lubrication, sizes 12-30F

Medical Equipment(32-51) Cont.

- [44] Bandage shears, 1
- [45] Extremity splints, adult, two each long and short
- [46] Extremity splints, pediatric, two each long and short

Rigid cervical collars, one each:

- [47] Pediatric
- [48] Small
- [49] Medium
- [50] Large
- [51] Restraints, arm and leg

Medical Supplies(52-83)

- [52] Trauma dressing, 6
- [53] Sterile gauze pads, 4" x 4", 20 each
- [54] Bandages, soft, roller, self adhering, 4" x 5 yd, 10 each
- [55] Vaseline gauze, 3" x 8", 2 each
- [56] Adhesive tape rolls, 2 each
- [57] Triangular bandages or slings, 5 each
- [58] Burn sheets, clean individually wrapped, 2 each
- [59] Sterile solution, 2000cc plastic bottles or bags
- [60] Aluminum foil or silver swaddler, 1 each with head cover
- [61] Obstetrical kit, sterile, 1 each
- [62] Cold packs, 3 each
- [63] Emesis basin, 1 each
- [64] Drinking water, 1 quart, sterile water may substitute
- [65] Ambulance emergency run reports, 10 each, with data as outlined in 515 Appendix E
- [66] Pillows, 2 each

- [67] Pillowcases, 2 each
- [68] Sheets, 2 each
- [69] Blankets, 2 each
- [70] CPR mask with safety valve, 1 each
- [71] Urinal, 1 each
- [72] Bedpan, 1 each
- [73] Nonporous disposable gloves
- [74] Impermeable red biohazard-labeled isolation bag
- [75] Face protection, mask/eye protection and/or field shields
- [76] Child/Infant car seat
- [77] Equipment/drug dosage sizing tape

OR

- [78] Pediatric equipment/drug age/weight chart
- [79] Pediatric trauma score reference
- [80] Poison control phone number
- [81] Plastic baby bottle w/nipple
- [82] Flashlight, 1 each, battery operated
- [83] Ambulance to hospital radio tested and working

Extrication Equipment(84-86)

- [84] Wrecking bar, 24", 1 each
- [85] Goggles
- [86] Fire extinguishers, 2 - 5lb ABC

General Observations(87-88)

- [87] Patient area clean
- [88] Equipment is secured in patient area
-

Inspector's Comments:
